

Pilot Point of Care C-Reactive Protein (CRP) Testing

Accreditation Status:	<i>currently not accredited to ISO 17043:2010</i>		
Date Scheme started:	2017		
Clinical Applicability:	Monitoring of the acute phase response		
Analytes:	C-Reactive Protein		
Units for Reporting:	mg/L		
Samples Distributed:	Liquid format. Normal and pathological human serum		
	Additional materials may be produced for specific recovery experiments by the addition of purified CRP to an analyte-free serum matrix		
Number of Distributions per year:	4		
Number of Samples per Distribution:	2		
Frequency of Distributions:	Currently seasonal as outlined in the Distribution Schedule		
Schedule of Analysis:	Data entry is via the web for the submission of results. Data analysis is commenced 14 days after sample dispatch. Late returns are accepted and will contribute to the laboratory's cumulative performance statistics		
Data Analysis:	All Laboratory Trimmed Mean (ALTM) with truncation at 2SD, 3SD, and CV%. Reports also show method specific statistics. Individual laboratory performance is expressed in terms of MRBIS, SDBIS, and MRVIS Chosen Coefficient of Variation for C-Reactive protein is 8%		
Performance Scoring:	MRVIS		
Criteria of Performance:	Laboratory performance is classified in terms of the MRVIS over a running analytical window of 4 Distributions (12 months)		
	Ideal	MRVIS	<50
	Good		50 - 100
	Adequate		101 - 200
	Poor		>200 or SDBIS >200
Persistent Poor Performance:	Defined as being in the Poor Performance category for two or more successive Distributions		