

**Bullous Dermatitis Antibodies**

<b>Accreditation Status:</b>	<b>UKAS Schedule of Accreditation</b>								
<b>Date Scheme started:</b>	1995								
<b>Clinical Applicability:</b>	Diagnosis of Bullous Dermatitis								
<b>Analytes:</b>	Dermatitis Basement Membrane and Desmosome Antibodies								
<b>Units for Reporting:</b>	Positive or Negative, U/mL, or titre as appropriate								
<b>Samples Distributed:</b>	Liquid format. Normal and pathological human serum								
<b>Number of Distributions per year:</b>	6								
<b>Number of Samples per Distribution:</b>	1								
<b>Frequency of Distributions:</b>	Every two months as outlined in the <b>Distribution Schedule</b>								
<b>Schedule of Analysis:</b>	<b>Data entry</b> is via the web for the submission of results. Data analysis is commenced 21 days after sample dispatch. Late returns are accepted and will contribute to the laboratory's cumulative performance statistics								
<b>Data Analysis:</b>	Qualitative responses are assessed by MI scoring in relation to the designated response								
<b>Performance Scoring:</b>	MI scoring and OMIS for all analytes for which the laboratory is registered								
<b>Criteria of Performance:</b>	Laboratory performance is assessed over a running analytical window of 6 Distributions (12 months)  The categories of performance are: <table> <thead> <tr> <th></th> <th><u>Total MIS</u></th> </tr> </thead> <tbody> <tr> <td>Good</td> <td>Zero</td> </tr> <tr> <td>Adequate</td> <td>1</td> </tr> <tr> <td>Poor</td> <td>&gt;1</td> </tr> </tbody> </table>		<u>Total MIS</u>	Good	Zero	Adequate	1	Poor	>1
	<u>Total MIS</u>								
Good	Zero								
Adequate	1								
Poor	>1								
<b>Persistent Poor Performance:</b>	Defined as being in the Poor Performance category for two or more successive Distributions								